



November 12, 2013

B. BRAUN MEDICAL INCORPORATED  
AVITUM  
DUANE MARTZ  
824 TWELFTH AVENUE  
BETHLEHEM PA 18018

Re: Description: HEALTH CARE PLUMBING APPLIANCE  
Manufacturer: B. BRAUN MEDICAL INCORPORATED  
Product Name: HEMODIALYSIS SYSTEM  
Model Number(s): DIALOG +  
Product File No: 20130309

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of November 2018.

This approval is contingent upon compliance with the following stipulation(s):

- This device does not require cross connection control measures on the influent water connections because the device contains an integral air gap that exceeds the requirements of ASME A112.1.2. The device contains a 1.25-inch air gap serving a 0.2-inch I.D. influent water tube (6.25 x pipe diameter).
- The wastewater from this device must discharge via an air gap that conforms to A112.1.2 to properly sized receptor. The materials used must be compatible with the temperature of the water being conveyed.
- All tubing/piping must be compatible with the temperature of the water being conveyed.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter  
Environmental Engineer - Plumbing Product Reviewer  
Department of Safety and Professional Services  
Division of Industry Services  
Bureau of Technical Services  
(608) 267-1401 Phone  
(608) 267-9723 Fax  
glen.schlueter@wi.gov E-mail